This policy has been adopted by UNC Health Care for its use in infection control. It is provided to you as information only.

Status ( Active ) PolicyStat ID ( 11346567

Origin	ation 0		Owner	Sherie Goldbach
UNC MEDICAL Appr CENTER Effe		05/2022	Policy Area	Infection Prevention
Last Rev	ctive 0 vised 0	)5/2022 )5/2022	Applicability	UNC Medical Center
Next Re	view 0	)5/2025		

#### **Cleaning And Disinfection of Non-Critical Items**

## I. Description

This policy is to provide UNC Medical Center health care personnel (HCP) guidance with the cleaning and disinfection of non-critical reusable patient care equipment, reusable patient care items and environmental surfaces.

## II. Rationale

An effective infection prevention and control policy is necessary to mitigate transmission of disease from direct or indirect contact with contaminated surfaces, patient equipment or patient care items.

## **III. Policy**

### **A. Spaulding Classification Scheme**

The Spaulding classification scheme is a rational approach to disinfection and sterilization of reusable patient care equipment and/or devices. Based on the degree of risk of infection involved in the use of items, the scheme divides reusable patient care items into three distinct categories: 1) critical; 2) semi-critical; 3) non-critical.

Critical Items: require sterilization	Enter/contact normally sterile tissue or the vascular system. Examples include but not limited to surgical instruments, certain catheters, implants, laparoscopes, arthroscopes.
Semi-critical Items: require at least high-level disinfection	Contact intact mucous membranes or non-intact skin. Examples include but not limited to some endoscopes, endocavitary probes, diaphragm fitting rings, laryngeal blades.

Non-critical Items: require<br/>low-level disinfectionContact intact skin but not mucous membranes. Examples<br/>include but not limited to bed pans, blood pressure cuffs, and<br/>stethoscopes.

- 1. For sterilization of critical items, refer to Infection Prevention policy: <u>Sterilization of Reusable</u> Patient Care Items
- 2. For High Level Disinfection of Semi-Critical items, refer to Infection Prevention policy: <u>High</u> Level Disinfection (HLD) – Manual reprocessing of reusable semi-critical medical devices

# B. Low Level Disinfection of non-critical reusable patient equipment, reusable patient care items and environmental surfaces.

- 1. Disinfectants
  - An EPA-registered disinfectant, for example Super Sani-Cloth<sup>®</sup>, Metriguard<sup>™</sup>, per manufacturer's instructions for use (MIFU) should be used for cleaning surfaces, reusable patient equipment, and reusable patient care items per MIFU.
  - b. In cases of suspected or confirmed diarrheal illness, items should be disinfected with a sporicidal disinfectant such as Sani-Cloth<sup>®</sup> bleach wipes per MIFU.
  - c. If MIFU do not provide cleaning and disinfection instructions, contact Infection Prevention for assistance.
- 2. Reusable non-critical patient care equipment or patient care items must be disinfected following the MIFU.
  - a. Follow MIFU for frequency of disinfection of non-critical patient care equipment or patient care items.
  - b. If MIFU do not specify a frequency of disinfection, all patient care equipment should be disinfected when visibly soiled, after use on a contact precautions or enteric contact precautions patient and on a regular basis (e.g., daily, weekly, etc.).

NOTE: Single use disposable patient care items may not be disinfected for reuse. Refer to Infection Prevention Policy: <u>Reuse of Single Use Devices</u>.

#### **C. Responsibilities**

- 1. Department Directors, Managers and/or designees
  - · Ensures surfaces, reusable patient care equipment and reusable patient care

items are cleaned and disinfected according to the manufacturer's instructions for use.

- 2. Infection Prevention Department
  - Consults on the cleaning and disinfection of surfaces, reusable patient equipment and reusable patient care items. Conducts rounding to assess compliance. Escalates concerns as needed and collaborates with unit or clinic leadership to ensure safe patient care.

## **IV. References**

CDC guideline for disinfection and sterilization in health-care facilities; Rutala WA, Weber DJ, and Healthcare Infection Control Advisory Committee (HICPAC). 2008

## **V. Responsible for Content**

Infection Prevention Department

## **VI. Related Policies**

Infection Prevention Policy: High-Level Disinfection (HLD) - Manual Reprocessing of Reusable Semi-Critical Medical Devices

Infection Prevention Policy: Reuse of Single Use Devices

Infection Prevention Policy: Sterilization of Reusable Patient Care Items

#### **Approval Signatures**

Step Description	Approver	Date
Policy Stat Administrator	Kimberly Novak-Jones: Nurse Educator	05/2022
	Thomas Ivester: CMO/VP Medical Affairs	05/2022
	Emily Vavalle: Dir Epidemiology	03/2022
	Sherie Goldbach: Project Coordinator	03/2022